



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

LabCor Portland

4321 SW Corbett Avenue, Suite A, Portland, OR 97239

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2005

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

Environmental and Chemical Testing *(As detailed in the supplement)*

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President/Operations Manager

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

Initial Accreditation Date:

June 28, 2017

Issue Date:

June 28, 2017

Expiration Date:

June 30, 2019

Accreditation No.:

90427

Certificate No.:

L17-276

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

LabCor Portland

4321 SW Corbett Avenue, Suite A, Portland, OR 97239
Contact Name: Kate March Phone: 206-781-0155

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Environmental, Chemical ^F	Wipes, Paint, Soil	Lead content	EPA 7000B	Detection Limit = 0.001 mg/kg
	Waste Water, Potable Water	Lead content	EPA 7010	Detection Limit = 0.001 mg/L
	Air	Lead content	NIOSH 7105	0.0005 mg/L to 0.01 mg/L
Environmental ^F	Air	Respirable Silica	NIOSH 7500	Range: 0.02 mg to 2 mg Detection Limit = 0.01 mg SiO ₂
		Asbestos (Direct Analysis)	ISO 10312	Detection limit calculated from requested analytical sensitivity
		Asbestos (Indirect Analysis)	ISO 13794	Detection limit calculated from requested analytical sensitivity
		Asbestos	NIOSH 7402	Detection limit calculated from requested analytical sensitivity
	Dust	Asbestos in Surface Dust	ASTM Dust D5755	Detection Limit = 1 000 str/mm ²

- The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location. Example: Outside Micrometer^F would mean that the laboratory performs this testing at its fixed location.